



CLEAN OIL
BRIGHT IDEAS

DIGI Field Kit EasySHIP

Oil Analysis Kit - Fuel, Hydraulic and Lube Oils

CJC™ Product Sheet

APPLICATION

The CJC™ DIGI Field Kit EasySHIP is an oil analysis kit used for measuring and monitoring **water in oil content, TBN, insolubles, viscosity and TAN** for industry, marine and off-shore equipment such as **diesel engines, gear boxes, compressors, hydraulics and turbines**. The CJC™ DIGI Field Kit EasySHIP is used extensively on oil types such as fuel, hydraulic and lube oil.

PROBLEMS IN OIL

Problems in oil can be related to corrosion, cavitation, instability (water), acidic combustion products (TBN), insolubles, viscosity, inorganic acids (TAN).

FUNCTION

The CJC™ DIGI Field Kit EasySHIP is specially provided for **on-site analysis**, which gives an easy and quick measurement of the oil condition.

An oil sample is taken from the system, entered into the DIGI Cell and mixed with reagents. The DIGI Cell is shaken for 3 min. and the result is shown on the graphic display of the DIGI Cell.

- Test time:** 3 min.
Nos. of tests: 50
Feature: Built-in memory for recording previous test result.

CAUTION

Please note that **TAN** is considered hazardous goods and shall be treated as such.
See Instr. Manual for Material Safety Data Sheet.

REAGENTS & SPARES

The DIGI Field Kit EasySHIP uses chemical reagents to test the oil. Replacement packs can be ordered as refill.

HOW TO ORDER

To order the DIGI Field Kit EasySHIP please use article no. in table below. Call your local C.C.JENSEN contact or enter our website.

PRODUCT DETAILS		
Article no.	Description	Content
FD35344	DIGI Field Kit EasySHIP (Complete suitcase)	See page 2
Replacement packs		
FD35347	EasySHIP Water in Oil Reagent Pack	Reagent A + EasySHIP Water in Oil Test Reagent
FD35291	ECON TAN Drop Test Kit	Reagent D + TAN
FD35293	TBN Reagent Pack	Reagent C
FD35294	Chromatography Papers	100 off




The Oil Analysis Kit
CJC™ DIGI Field Kit EasySHIP
Contains the DIGI Cell, Reagents, Spare Parts and Instruction Manual.



The DIGI Cell

- Shows the results on the graphic display
- Offers on-screen user guide
- The DIGI Cell is used to monitor water in oil and TBN



TECHNICAL DATA		
AREA	RANGE	ACCURACY
Water	0.02 - 1%, 200 - 10,000 ppm, 0-10%, 0-20%	Typically ± (1000 ppm) or ± 1% (on 0-10% range)
TBN	5 - 55	Typically within ± 5%
Insolubles	qualitative	
Viscosity	go / no go	
TAN	0-3 / 0-6 mg KOH TAN	Typically within ± 0.2 TAN



CLEAN OIL
BRIGHT IDEAS

DIGI Field Kit EasySHIP

Oil Analysis Kit - Fuel, Hydraulic and Lube Oils

CJC™ Product Sheet

CONTENT IN SUITCASE

DIGI Field Kit EasySHIP	
Reagent A	500 ml
EasySHIP Water in Oil Test Reagent	50 sachets
Reagent C	500 ml
Reagent D	TAN
TAN titrant	30 ml, KS368
DIGI Cell	
Agitators	plastic rings for Cell, 3 off
Syringes	disposable 10 ml
Scissors	stainless steel, blue
Tweezer	disposable 3.5 IN
Plastic gloves	disposable pack
Plastic beaker	100 ml
Visco stick	blue
Chromatography papers	100 off
Plastic titration jar	with cap, 60 ml
DIGI instr. manual	
Aluminium case	large

PRODUCT SAFETY SHEET

Read this BEFORE you use the Equipment

Please ensure:

- The operator has read and understood the instruction manual.
- If the equipment is connected to a main power supply, ensure that the equipment is properly earthed and the correct supply voltage used.
- The working area is well ventilated and illuminated.
- The equipment is not subjected to damp or liquid spray.
- The equipment is never used in a hazardous environment.
- The operator has read and understood all Materials Safety Data Sheets before use (please see instruction manual), and is fully familiar with any hazards and necessary precautions.
- All reagents are sealed tightly after use and cannot leak.
- All equipment is emptied of reagents and oil after use and thoroughly cleaned.
- Equipment is only used for it's intended use, using it any other way may impair any safety protection.
- No smoking during use of the equipment or when handling any of the reagents.
- Use protective goggles and gloves.

Use of Pressure Cells

Some tests require the use of pressure cells. They are safety tested at pressures 10 times normal use.

- Do not exceed the recommended volumes of reagents or test oils in the cell.
- The cell is designed to gently vent excess pressure when opened. Hold the cell upright, away from the body and slacken the cap slowly to release the pressure and prevent any liquid being sprayed out.